

Hold-n-Place® Catheter Securement

Application Instructions

- For IV and arterial catheters the Luer connection should be attached to the inserted catheter
- Clean & Prep the skin per hospital protocol.
- On some patients hair removal or skin prep may be required for better adhesion.
- Make sure the skin is clean and dry before applying the securement device.
- 5. Pull the top adhesive PULL tab back to the perforations on the two strips. (A)
- Slide the securement device under the catheter/sheath positioning where the catheter tube joins the hub is centered over the extended point in the center of the base. (B)
- 7. Apply gentle pressure to the top of the hub to secure in the exposed adhesive. (B)
- Hold the securement device and the catheter in position while pulling one bottom release liner and press adhesive to skin, repeat for other side. (C)
- Remove the PULL tab completely to expose the adhesive strips. (D)
- 10. Secure one adhesive strip over and across the catheter hub applying gentle pressure to the top side of the adhesive strip, then repeat with the other adhesive strip to achieve a chevron securement. (E)
- 11. Press down on securement device to ensure better adhesion
- 12. Apply a transparent dressing over the insertion site. You can place the transparent dressing over the entire securement device, however, only mid-way covering the securement device is recommended to ease dressing changes.
- 13. Follow hospital protocol for dressing change interval. If soiled or saturated in fluid, replace securement device.

Tape-Strips are included to help stabilize line or tube while applying the securement device.

Removal

Alcohol swab may be used to help with removal.

Ordering Information

849	Hold-n-Place IV, arterial and mid-line	50/00
	catheter securement with transparent dressing	

850	Hold-n-Place IV, arterial and mid-line	50/box
	catheter securement	

Catrieter Securement	
 	50 A

	catheter securement with Prevantics® in	
	an IV Start Kit	

852	Hold-n-Place CVC, PICC and arterial	50/box
	sheath securement	

Hold-n-Place CVC, PICC and arterial 50/hox 853 sheath securement with large transparent dressing

For single patient use. Disposable.

For an application video please visit www.dalemed.com

STERILE R

Dale medical products are available from your preferred hospital/medical supply dealer.

Not made with natural rubber latex.

Patent Pending

Dale® & Hold-n-Place® are registered trademarks of Dale Medical Products, Inc. in the USA & EU.

TEL.: +1-508-695-9316 www.dalemed.com



MDSS GmbH REP

Schiffgraben 41 D-30175 Hannover, Germany

946001 REV F

©Copyright 2020 English









